



QUALITY CONTROL RANGES
Milliplex MAP® Rat Cardiac Injury
Magnetic Bead Panel 1
Kit Catalog # RCI1MAG-87K
Control Catalog # RCI1-6087-1
Lot# RCI1-104 and RCI1-204

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
Troponin I	Control I	758 – 1574	pg/mL
	Control II	8924 – 18534	pg/mL
Troponin T	Control I	436 – 1036	pg/mL
	Control II	6314 – 13113	pg/mL
CKM	Control I	120 – 274	pg/mL
	Control II	1496 – 3444	pg/mL
FABP3	Control I	253 – 581	pg/mL
	Control II	3350 – 7475	pg/mL
Follistatin-like 1	Control I	549 – 1140	pg/mL
	Control II	6209 – 12896	pg/mL
MYL3	Control I	27 – 56	pg/mL
	Control II	336 – 699	pg/mL
TIMP1	Control I	303 – 630	pg/mL
	Control II	3612 – 7502	pg/mL

Note: The Quality Control Ranges are generated using serum matrix provided in the kit. Quality Control values in culture media are not tested.

RCI1-6087-1: RCI1-104.204 (04 February 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.