

HybriScan® D Lactobac

**Rapid test system for qualitative detection of
Lactobacteria in non-alcoholic drinks**

Product-No.: 59744



Contact information:**HybriScan® - Rapid Test System (R&D)**Dr. Helmut Maucher
Phone: (+49) – 3494 – 6364 15
e-mail: contact@scanbec.de**Sales Organisations****Argentina**SIGMA-ALDRICH DE
ARGENTINA S.A.
Free Tel: 0810 888 7446
Tel: (+54) 11 4556 1472
Fax: (+54) 11 4552 1698**Australia**SIGMA-ALDRICH PTY LTD.
Free Tel: 1800 800 097
Free Fax: 1800 800 096
Tel: (+61) 2 9841 0555
Fax: (+61) 2 9841 0500**Austria**SIGMA-ALDRICH HANDELS
GmbH
Tel: (+43) 1 605 81 10
Fax: (+43) 1 605 81 20**Belgium**SIGMA-ALDRICH NV/SA.
Free Tel: 0800 14747
Free Fax: 0800 14745
Tel: (+32) 3 899 13 01
Fax: (+32) 3 899 13 11**Brazil**SIGMA-ALDRICH BRASIL LTDA.
Free Tel: 0800 701 7425
Tel: (+55) 11 3732 3100
Fax: (+55) 11 5522 9895**Canada**SIGMA-ALDRICH CANADA LTD.
Free Tel: 1800 565 1400
Free Fax: 1800 265 3858
Tel: (+1) 905 829 9500
Fax: (+1) 905 829 9292**China**SIGMA-ALDRICH (SHANGHAI)
TRADING CO. LTD.
Free Tel: 800 819 3336
Tel: (+86) 21 6141 5566
Fax: (+86) 21 6141 5567**Czech Republic**SIGMA-ALDRICH S.R.O.
Tel: (+420) 246 003 200
Fax: (+420) 246 003 291**Denmark**SIGMA-ALDRICH DENMARK
A/S
Tel: (+45) 43 56 59 10
Fax: (+45) 43 56 59 05**Finland**SIGMA-ALDRICH FINLAND OY
Tel: (+358) 9 350 9250
Fax: (+358) 9 350 92555**France**SIGMA-ALDRICH CHIMIE S.à.r.l.
Free Tel: 0800 211 408
Free Fax: 0800 031 052
Tel: (+33) 474 82 28 00
Fax: (+33) 474 95 68 08**Germany**SIGMA-ALDRICH CHEMIE
GmbH
Free Tel: 0800 51 55 000
Free Fax: 0800 64 90 000
Tel: (+49) 89 6513 0
Fax: (+49) 89 6513 1160**Greece**SIGMA-ALDRICH (O.M.) LTD.
Tel: (+30) 210 994 8010
Fax: (+30) 210 994 3831**Hungary**SIGMA-ALDRICH Kft
Ingyenes zöld telefon: 06 80 355
355
Ingyenes zöld fax: 06 80 344 344
Tel: (+36) 1 235 9055
Fax: (+36) 1 235 9050**India**SIGMA-ALDRICH CHEMICALS
PRIVATE LIMITED
Telephone
Bangalore: (+91) 80 6621 9600
New Delhi: (+91) 11 4165 4255
Mumbai: (+91) 22 2570 2364
Hyderabad: (+91) 40 6684 5488
Fax
Bangalore: (+91) 80 6621 9650
New Delhi: (+91) 11 4165 4266
Mumbai: (+91) 22 2579 7589
Hyderabad: (+91) 40 6684 5466**Ireland**SIGMA-ALDRICH IRELAND
LTD.
Free Tel: 1800 200 888
Free Fax: 1800 600 222
Tel: (+353) 1 404 1900
Fax: (+353) 1 404 1910**Israel**SIGMA-ALDRICH ISRAEL LTD.
Free Tel: 1 800 70 2222
Tel: (+972) 8 948 4100
Fax: (+972) 8 948 4200**Italy**SIGMA-ALDRICH S.r.l.
Numero Verde: 800 827018
Tel: (+39) 02 3341 7310
Fax: (+39) 02 3801 0737**Japan**SIGMA-ALDRICH JAPAN K.K.
Tokyo Tel: (+81) 3 5796 7300
Tokyo Fax: (+81) 3 5796 7315**Korea**SIGMA-ALDRICH KOREA
Free Tel: (+82) 80 023 7111
Free Fax: (+82) 80 023 8111
Tel: (+82) 31 329 9000
Fax: (+82) 31 329 9090**Malaysia**SIGMA-ALDRICH (M) SDN. BHD
Tel: (+60) 3 5635 3321
Fax: (+60) 3 5635 4116**Mexico**SIGMA-ALDRICH QUÍMICA,
S.A. de C.V.
Free Tel: 01 800 007 5300
Free Fax: 01 800 712 9920
Tel: 52 722 276 1600
Fax: 52 722 276 1601**The Netherlands**SIGMA-ALDRICH CHEMIE BV
Free Tel: 0800 022 9088
Free Fax: 0800 022 9089
Tel: (+31) 78 620 5411
Fax: (+31) 78 620 5421**New Zealand**SIGMA-ALDRICH NEW
ZEALAND LTD.
Free Tel: 0800 936 666
Free Fax: 0800 937 777
Tel: (+61) 2 9841 0555
Fax: (+61) 2 9841 0500**Norway**SIGMA-ALDRICH NORWAY AS
Tel: (+47) 23 17 60 60
Fax: (+47) 23 17 60 50**Poland**SIGMA-ALDRICH Sp. z o.o.
Tel: (+48) 61 829 01 00
Fax: (+48) 61 829 01 20**Portugal**SIGMA-ALDRICH QUÍMICA,
S.A.
Free Tel: 800 202 180
Free Fax: 800 202 178
Tel: (+351) 21 924 2555
Fax: (+351) 21 924 2610**Russia**SIGMA-ALDRICH RUS, LLC
Tel: +7 (495) 621 6037
Fax: +7 (495) 621 5923**Singapore**SIGMA-ALDRICH PTE. LTD.
Tel: (+65) 6779 1200
Fax: (+65) 6779 1822**South Africa**SIGMA-ALDRICH
SOUTH AFRICA (PTY) LTD.
Free Tel: 0800 1100 75
Free Fax: 0800 1100 79
Tel: (+27) 11 979 1188
Fax: (+27) 11 979 1119**Spain**SIGMA-ALDRICH QUÍMICA,
S.A.
Free Tel: 900 101 376
Free Fax: 900 102 028
Tel: (+34) 91 661 99 77
Fax: (+34) 91 661 96 42**Sweden**SIGMA-ALDRICH SWEDEN AB
Tel: (+46) 8 742 4200
Fax: (+46) 8 742 4243**Switzerland**SIGMA-ALDRICH CHEMIE
GmbH
Free Tel: 0800 80 00 80
Free Fax: 0800 80 00 81
Tel: (+41) 81 755 2828
Fax: (+41) 81 755 2815**United Kingdom**SIGMA-ALDRICH COMPANY
LTD.
Free Tel: 0800 717 181
Free Fax: 0800 378 785
Tel: (+44) 1747 833 000
Fax: (+44) 1747 833 313
SAFC (UK) Free Tel: 0800 71 71
17**United States**SIGMA-ALDRICH
P.O. Box 14508
St. Louis, Missouri 63178
Toll-Free: 800 325 3010
Toll-Free Fax: 800 325 5052
Call Collect: (+1) 314 771 5750
Tel: (+1) 314 771 5765
Fax: (+1) 314 771 5757**Internet**sigma-aldrich.com**Technical Service:**flukatec@sial.com**Product Specifications****Cat. No.:** 59744
Number of tests: 96 tests, incl. standard series
Storage: 4 – 8°C, 12 month
Test duration: approx. 2-2.5 hours (after pre-enrichment)
Sensitivity: 1 -10 CFU/L (after pre-enrichment)
Specificity: bacteria of the genera *Lactobacillus* and *Pediococcus*

HybriScan® D Lactobac–Test Protocol

Working Principle

HybriScan® D Lactobac is a rapid molecular test system for the detection of bacteria of the genera *Lactobacillus* and *Pediococcus* in only one test. HybriScan® D Lactobac is based on the detection of target molecules from the microorganism of interest by means of specific capture and detection probes in a so-called sandwich hybridization. The target molecules of the beer contaminants contained in the sample are captured in a specific microtiter binding plate. All other unbound sample components are removed by several washing steps, so that only beer-spoiling microbes are detected in a highly specific way. In addition to the capture probe, a detection probe is coupled to the target molecule. An enzyme is attached afterwards in a subsequent incubation step. After several washing steps, reaction with a colour substrate gives a blue colouration, which changes to yellow after the addition of a stop solution. The yellow colour enables highly sensitive photometric measurement at 450 nm. Comparison is made with the standard solutions contained in the test kit.

Technical Notes

After starting the test procedure, perform each of the following steps without interruption and within the given time limits:

For each sample use an individual single-use pipette tip to avoid cross-contamination.

Close bottles immediately after use and store them at the temperatures specified on the labels. Do not interchange caps and bottles.

Samples and standards should be tested together for more accurate results.

Do not mix or replace components from test kits of different charges.

Incubation at room temperature refers to a laboratory temperature of 20 to 25°C.

Do not use the test kit after the expiration date listed on the package.

Safety

All reagents contained in the test kit are for *in vitro* use only.

Test solution D contains formamide. Avoid contact with eyes, skin and the respiratory system. In event of contact with eyes or skin, rinse immediately with plenty of water. If the reagent is inhaled, immediately remove the individual to fresh air and seek medical attention.

Stop solution H contains 1 N sulfuric acid. Avoid contact with eyes and skin. In the event of contact with eyes and skin rinse immediately with plenty of water.

Handling of the kit components and disposal of waste should be performed according to standard laboratory safety guidelines.

Reagents and Storage Conditions

The reagents contained in the test kit are sufficient for at least 96 tests, including 6 standard series. The kit components should be stored between +2 to +8°C as indicated on the labels. Do not freeze the test kit components!

Kit components:

1. Microwell plate , ready to use, 96 wells	1
2. Binding plate , ready to use, 96 wells	1
3. Standards 1 – 4^{a)} (white screw caps); blank and different concentrations of synthetic RNA as positive control and for preparation of a calibration curve for quantitative measurements	0.2 mL each
4. Lysis Reagent A (red screw cap), ready to use	1.2 mL
5. Lysis Buffer B^{a)} (red cap), ready to use	4.5 mL
6. Lysis Buffer C^{a)} (red cap), ready to use	5.5 mL
7. Test Solution D (yellow cap), ready to use	4.5 mL
8. Washing Solution E^{b)} (blue cap), ready to use	90 mL
9. Enzyme Solution F (green screw cap), dilute a suitable amount 1:100 with Washing Solution E before use	0.120 mL
10. Substrate Solution G^{b)} (green cap), ready to use	10 mL
11. Stop Solution H (green cap) 1 N sulfuric acid, ready to use	5 mL
12. Glass beads (colourless cap), sterile, ready to use	4 mL

^{a)}Components contain SDS, which precipitates at lower temperatures. Equilibrate to room temperature before use.

^{b)}Equilibrate to room temperature before use.

Additional equipment and materials (required, not supplied with kit)

- Centrifuge for microreaction tubes (1.5 and 2 mL)
- Thermoshaker for microreaction tubes and microwell plates
- Vacuum filtration unit
- 3 Pipettes (2–20 µL, 20–200 µL, 200–1000 µL) with corresponding tips; optional 8-channel pipette (20–200 µL)
- Microwell plate-photometer
- Enrichment medium, incubator
- Microreaction tubes (2 mL), cultivation tubes (12 mL), reagent-reservoirs, membrane filter discs (0.45 µm)

Specificity

Following species of microorganisms can be detected with the HybriScan® **D** *Lactobac*-Kit:

- **Genus *Lactobacillus*:**
 - Lactobacillus acidophilus*
 - Lactobacillus brevis*
 - Lactobacillus brevisimilis*
 - Lactobacillus buchneri*
 - Lactobacillus casei*
 - Lactobacillus collinoides*
 - Lactobacillus coryniformis*
 - Lactobacillus curvatus*
 - Lactobacillus fermentum*
 - Lactobacillus fructivorans*
 - Lactobacillus lindneri*
 - Lactobacillus malefermentans*
 - Lactobacillus parabuchneri (frigidus)*
 - Lactobacillus paracasei*
 - Lactobacillus paraplantarum*
 - Lactobacillus plantarum*
 - Lactobacillus rhamnosus*

- **Genus *Pediococcus*:**
 - Pediococcus acidilactici*
 - Pediococcus claussenii*
 - Pediococcus damnosus*
 - Pediococcus inopinatus*
 - Pediococcus parvulus*
 - Pediococcus pentosaceus*

Test protocol

(1) Sample preparation

Filter 100–1,000 mL of sample (vacuum filtration unit; membrane filter disc; 0.45 µm pore size). After filtration incubate the filter disc in 5 mL of enrichment medium (e.g. MRS-Broth, 28°C, 24 h)

Transfer a 2 mL aliquot from the pre-cultivation tube via pipette to a 2 mL microreaction tube that contains a spatula-tip amount of glass beads. Centrifuge the samples for 2 minutes at maximum speed of 13,000 rpm. Remove the supernatant carefully with a pipette.

Note:

Avoid strong shaking after centrifugation to avoid resuspending the bacteria pellet. Centrifuge a second time if necessary. Samples with high yeast content should be centrifuged for 1 minute at 1,000 rpm to allow the yeast to settle. Take 1 mL of the supernatant and proceed with step 1. To increase the sensitivity, start with 2 x 2 mL sample and combine the supernatants.

(2) Cell lysis

Add 40 µL of **Lysis Buffer B** (bottle with red cap) to the cell pellet and 10 µL of **Lysis Reagent A*** (microreaction tube with red screw cap), mix well and incubate for 15 minutes at 37°C in a thermoshaker. Next, add 50 µL of **Lysis Buffer C** (bottle with red cap). Incubate again for 15 minutes at 37°C with shaking at 1,400 rpm in the thermoshaker. Centrifuge the samples for 10 minutes at 13,000 rpm using a microcentrifuge. Use 10 µL of this supernatant in protocol step 3 (hybridization).

***Note:** In the case of a large number of samples prepare a Master Mix of **Lysis Reagent A** and **Lysis Buffer B** before use. Pipette 50 µL of the Master Mix to each cell pellet.

Preparation for subsequent steps:

Change the top of the thermoshaker and fix the manifold for microwell plates. Set the temperature to 50°C and shaking speed to 500 rpm. For quantitative analysis of beer-spoiling microorganisms we recommend a repeat determination of the 4 standard solutions (microreaction tubes with white screw cap).

To each well of standards (repeat determinations, e.g. A1-H1) and samples add 45 µL of Test Solution D (bottle with yellow cap). Incubate the plate at 50°C for 5 minutes in the thermoshaker.

(3) Hybridization

Add 10 µL of **Standard 1** into the wells A1 and B1; 10 µL of **Standard 2** into C1 and D1; 10 µL of **Standard 3** into E1 and F1 and 10 µL of **Standard 4** into G1 and H1. Apply 10 µL of each sample (supernatant from step 2) into the respective well position. Cover the microwell plate with a lid and incubate it in the thermoshaker for 10 minutes at 50°C and 500 rpm.

Note:

When adding the standards and samples, to avoid cooling do not remove the microwell plate from the thermoshaker.

The supernatant from step 2 can be stored at -20°C for future use.

(4) Coupling to the binding plate

Transfer 50 µL of the reaction mixes from each well to the corresponding wells of the binding plate and shake for 10 minutes at 50°C and 500 rpm in the thermoshaker.

Note:

Unused stripes of the plate should be stored in the sealed original packing at 4 to 8 °C.

Preparation for subsequent steps:

The **Enzyme Solution F-Washing Solution E** 1:100 dilution must be prepared immediately before use. It cannot be stored. Prepare only the amount needed for the test, e. g. for 16 reactions combine 1700 µL **Washing Solution E** and 17 µL **Enzyme Solution F**.

Note:

Briefly spin down enzyme solution F prior use to collect the liquid at the bottom of the tube.

(5) Enzymatic reaction

Discard the liquid from each well by inverting and gently tapping of the plate on an absorbent layer. Set the temperature to 25°C. Add 200µL **Washing Solution E** (bottle with blue cap) and incubate for 2 minutes at room temperature. Discard the liquid. Pipette 100 µL of the diluted **Enzyme Solution F**, prepared as described above "preparation for subsequent steps", to each well. Cover the binding plate with a lid and incubate it in the thermoshaker for 10 minutes at 25°C and 500 rpm.

(6) Washing

Discard the liquid from each well. Add 200 µL of **Washing Solution E** (bottle with blue cap) to each well and incubate the microplate (with lid) for 1 minute at 25°C and 500 rpm in the thermoshaker. Repeat washing each well once.

Preparation for subsequent steps:

Switch on computer and the microplate reader.

(7) Substrate Reaction

After discarding the Washing Solution from the second wash step, add 100 µL of **Substrate Solution G** (bottle with green cap) to each well. Cover the microplate with a lid and incubate it in a thermoshaker for 10 minutes at 25°C and 500 rpm. Stop the reaction by adding 50 µL of **Stop Solution H** (bottle with green cap) to each well. The addition of acid creates a yellow colour change. Mix shortly (10 sec, 500 rpm) in the thermoshaker and remove air bubbles, if present.

Note:

For qualitative analysis results can be measured by visual inspection. Compared to the blanks (A1, B1), which should be colourless, a blue colour change indicates contamination of the sample.

Data analysis using the HybriScan®-Software**(8) Signal read-out and data analysis using the HybriScan®-Software**

Start the reader and open the HybriScan®-Software. Insert the microwell plate into the reader, with position A1 rear left. Click button **Start Measurement** and choose the appropriate test. Click **Start**. The instrument measures the absorbance of any well. Results are saved automatically in the database. Alternatively, results of tests saved as text file can be opened by choosing option **File/Open** in the main menu.

The graphical layout of the results corresponds to the position of the sample in the microwell plate. By clicking on each position, you can enter the name of the sample. In the lower pane of the graphical user interface the user name and other comments can be inserted. For each position you can hide or unhide parameters by choosing the button **View**.

Click the button **Report**. In the upper pane of the window the regression line calculated from the four standards is displayed. Values of standards are represented by red dots. They should be close to the regression curve (blue line). Your results can be printed by choosing the option **Print Report**. For post processing (like in Microsoft Excel) results can be exported by choosing the option **Export Report**.

(9) Data interpretation

Based on the absorbance measurement, HybriScan®-Software will help to evaluate the analyzed samples. The quantitative analysis is performed automatically on the basis of the standard values (A1 to H1). For each well the corresponding number of cells are calculated and indicated by the software

numerically and with a colour code. The change of green to red enables a fast visual recognition of contaminated samples.

The cell number calculated by the HybriScan®-Software reflects the amount of cell equivalents in the 10 µL sample introduced to the test solution. If you started with 2mL culture, you have to multiply the calculated cell number by 5 to get the amount of cells per mL.

Furthermore, the mean values of optical density of the negative control (standard 1) are used to evaluate, if a sample is positive, questionable or negative.

The HybriScan®-Software enables both, graphical and tabular representation of the measurement data of the respective samples. Optical density and calculated numbers of cells are displayed by the software tool.

Data analysis without the HybriScan®-Software

(10) Signal read-out using VIS-photometer

Start the reader and open the photometer Software. Insert the microwell plate into the reader, with position A1 rear left. Start the measurement. The instrument measures the absorbance of any position at 450 nm.

(11) Internal control

We recommend using all four standards as internal controls for the test procedure. If the analysis is performed correctly your measured data will be near the regression line. We recommend this method especially for unexperienced user.

For linear regression set standard1 = 0, standard2 = 4, standard3 = 8 and standard4 = 16 at the abscissa (x-coordinate). The straight line must subtend the y-axis at S1 = 0.

Regression line offers only verification of analysis-quality; a quantification is not possible.

For experienced user it is sufficient to use only standard S1 and S3. These standards are necessary for a qualitative data analysis.

(12) Qualitative analysis

For the measurement to be valid, the quotient of the mean value of the positive control (S3) divided by the mean value of the negative control (S1) must be greater than 4.0.

Evaluation of the samples is performed using the following formula:

$$\text{Sample OD\%} = \frac{\text{OD}_{\text{Sample}} - \text{MV OD}_{\text{NC}}}{\text{MV OD}_{\text{PC}} - \text{MV OD}_{\text{NC}}} \times 72.1 \text{OD\%}$$

MV mean value
 PC positive control (S3)
 NC negative control (S1)

Sample OD% values are used to evaluate the sample status:

Samples with OD% values under 10 are considered negative.

Samples with OD% values from 10 to < 20 are considered questionable.

Samples with OD% values ≥ 20 are considered positive.

(13) Quantitative analysis

For quantitative analysis (semi - quantitative analysis) use following formula to calculate cell count of your samples:

$$\text{Cell count}_{\text{Sample}} = \frac{\text{OD}_{\text{Sample}} \times \text{Cell count}_{\text{Standard}}}{\text{MV OD}_{\text{Standard}}}$$

MV mean value

For calculating the cell count, choose a standard, which has a OD next to sample – OD.

Cell count – standards:

S1 = 0

S2 = 4000

S3 = 8000

S4 = 16000

Note:

Quantitative analysis is only possible, if sensitivity was reached (see product specifications), your measured data are near the regression line and no pre-enrichment was performed!

Short Protocol

1. Filter 100–1,000 mL of sample (vacuum filtration unit; membrane filter disc; 0.45 µm pore size)
2. After filtration incubate the filter disc in 5 mL of enrichment medium
3. Remove 2 mL of sample from the enrichment medium, add glass beads, centrifuge (13,000 rpm, 2 minutes) and discard the supernatant
4. Add 40 µL of **Lysis Buffer B** (red cap) to the pellet and add 10 µL of **Lysis Reagent A** (red screw cap); mix and incubate for 15 minutes at 37°C in a thermoshaker
5. Add 50 µL of **Lysis Buffer C** (red cap) and incubate for 15 minutes at 37°C and 1,400 rpm in the thermoshaker
6. Centrifuge for 10 min at 13,000 rpm
7. Pipette 45 µL of **Test Solution D** (yellow cap) per sample (including the standards) into the wells of a microplate and pre-incubate for at least 5 minutes at 50°C and 500 rpm in the thermoshaker
8. Add 10 µL of the supernatant from step 6 to each well (row A1–H1 is reserved for the respective standards); cover the microwell plate with a lid and incubate for 10 min at 50°C and 500 rpm in the thermoshaker
9. Transfer 50 µL of reaction mixes to the binding plate and incubate for 10 min at 50°C and 500 rpm in a thermoshaker
10. Discard all liquid and wash the plate with 200 µL **Washing Solution E** (blue cap), discard Washing Solution
11. Dilute a suitable amount of **Enzyme Solution F** (green screw cap) 1:100 with **Washing Solution E** (blue cap) and add 100 µL of the mixture to each well of the binding plate; cover the plate with a lid and incubate for 10 minutes at 25°C and 500 rpm in the thermoshaker
12. Discard all liquid and add 200 µL of **Washing Solution E** (blue cap) to each well and incubate for 1 minute at room temperature and 500 rpm in the thermoshaker; repeat the washing step once
13. Discard all liquid and add 100 µL **Substrate Solution G** (green cap) per sample to the wells of the binding plate; cover the plate with a lid and incubate for 5-15 minutes at 25°C and 500 rpm in the thermoshaker
14. Add 50 µL **Stop Solution H** (green cap) to each well
15. Place the Plate in a microplate reader and measure the optical density in each well at 450 nm; perform data analysis

Overview of the HybriScan® D Lactobac procedure:



1. Sample filtration
(100 – 1000 ml,
15 min)



2. Enrichment culture, optional
(MRS-Broth, 24h)



3. Cell lysis
(2 ml sample, 13.000 rpm;
37°C, 45-60 min)



4. HybriScan® test solution
(Forming of "sandwich complexes"
between specific probes an the
sample , 10 min)



5. Immobilisation
(Binding of the "sandwiches"
to the binding plate, 10 min)




6. Enzyme coupling
(Coupling of an enzyme to the
"sandwiches", 10 min)



7. Washing
(Removal of unbound
components, 2x 1 min)



8. Signal read out
(Colour reaction and signal
read out, 15 min)



9. Test analysis
(Analysis of the data, 10 min)

Advantages

- Rapid, sensitive, reliable
- Specific for living cells
- Time saving of 2 to 4 days in comparison to cultivation based assays
- Easy to handle
- Minimal sample preparation
- High sample throughput using 96-well microplates