

Benchmark Scientific Inc.  
116 Corporate Blvd.  
South Plainfield, NJ 07080 USA

## DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

**Device Name:** BenchMixer™ Vortex mixer and Mortexer™ Vortex mixer

**Device Model Number(s):** BV1000-E and BV1005-E and Private label versions of these devices.

**Relevant EC Directives:** Low Voltage Directive 2006/95/EC  
EMC Directive 2004/108/EC

**Harmonized Standards:** EN 61010-1:01  
EN 61010-2-101:02  
EN 61010-2-051:95  
EN 55011:98/A1:99/A2:02  
EN 61000-3-2:00+A2:05  
EN 61000-3-3:95+A1:01+A2:05  
EN 61000-4-2:95 +A1:98 +A2:01  
EN 61000-4-3:02 +A1:02  
EN 61000-4-4:04  
EN 61000-4-5:95 +A1:01  
EN 61000-4-6:96 +A1:01  
EN 61000-4-8:93 +A1:01  
EN 61000-4-11:04  
EN 61326-1:06  
EN 61326-2-6:06

**Manufacturer's Name:** Benchmark Scientific, Inc.  
PO Box 709  
Edison, NJ 08818

**Date:** November 01, 2010



Marketing Manager